

Human HNP1-3

HK317
Edition 07-15

**ELISA KIT
PRODUCT INFORMATION & MANUAL**

Read carefully prior to starting procedures!
For use in laboratory research only
Not for clinical or diagnostic use



Note that this user protocol is not lot-specific and is representative for the current specifications of this product. Please consult the vial label and the Certificate of Analysis for information on specific lots. Also note that shipping conditions may differ from storage conditions.

For research use only. Not for use in or on humans or animals or for diagnostics. It is the responsibility of the user to comply with all local/state and federal rules in the use of this product. Hycult Biotech is not responsible for any patent infringements that might result from the use or derivation of this product.

TABLE OF CONTENTS

	Page
1. Intended use.....	2
2. Introduction.....	2
3. Kit features	2
4. Protocol overview	3
5. Kit components and storage instructions	4
6. Warnings and precautions	5
7. Sample preparation	6
8. Reagent preparation	8
9. ELISA protocol.....	9
10. Interpretation of results	10
11. Technical hints.....	10
12. Quality control.....	11
13. Performance Characteristics.....	11
14. Troubleshooting	12
15. References	13

1. INTENDED USE

The human HNP1-3 (Neutrophil Defensins) ELISA kit is to be used for the *in vitro* quantitative determination of human HNP1-3 in BALF, cell culture supernatant, cervicovaginal secretions, feces, plasma, saliva, serum and sputum samples. This kit is intended for laboratory research use only and is not for use in diagnostic or therapeutic procedures.

The analysis should be performed by trained laboratory professionals.

2. INTRODUCTION

Human neutrophil defensins (alpha-defensins, HNP1-3) belong to the family of cationic trisulfide-containing microbicidal peptides. Besides microbicidal, the peptides exert chemotactic, immunomodulating and cytotoxic activity and participate in host defense and inflammation. Azurophilic granules of neutrophils contain human Neutrophil Peptide (HNP)-1-4 which are highly homologous. The three principal human defensins, HNP1-3, are unique to neutrophils and account for about 99 percent of the total defensin content of these cells. Measured amount of defensins is 3-5 mg per million human neutrophils.

Activation of neutrophils leads to rapid release of defensins. Thus, only one cell type, neutrophils, may be the source of HNP1-3 measured in plasma and other body fluids during infection and inflammation. In normal plasma low levels of HNP1-3 are present ranging from undetectable level to 50-100 ng/ml, while in septic conditions the levels of HNP1-3 might be elevated to 10 mg/ml and even more. Activation of neutrophils in blood as occurs during clotting, as well as long storage of anticoagulated blood leads to a release of HNP, thus careful plasma sampling is important for possible detection of HNP. Defensins are relatively resistant to proteolysis, low pH and boiling, but have a tendency to bind to a variety of materials, including plastic and proteins.

3. KIT FEATURES

- Working time of 3½ hours.
- Minimum concentration which can be measured is 156 pg/ml.
- Measurable concentration range of 156 to 10,000 pg/ml.
- Working volume of 100 µl/well.

Cross-reactivity


Potential cross-reacting proteins detected in the human HNP1-3 ELISA:

Cross reactant	Reactivity
Bovine	No
Dog	No
Goat	No
Horse	No
Mouse	No
Rabbit	No
Rat	No
Rhesus monkey	Strong
Pig	No
Sheep	No

Table 1

Cross-reactivity for other species or proteins/peptides has not been tested.

4. PROTOCOL OVERVIEW



- The human HNP1-3 ELISA is a ready-to-use solid-phase enzyme-linked immunosorbent assay based on the sandwich principle with a working time of 3½ hours.
- The efficient format of a plate with twelve disposable 8-well strips allows free choice of batch size for the assay.
- Samples and standards are incubated in micro titer wells coated with antibodies recognizing human HNP1-3.
- Biotinylated tracer antibody will bind to captured human HNP1-3.
- Streptavidin-peroxidase conjugate will bind to the biotinylated tracer antibody.
- Streptavidin-peroxidase conjugate will react with the substrate, tetramethylbenzidine (TMB).
- The enzyme reaction is stopped by the addition of oxalic acid.
- The absorbance at 450 nm is measured with a spectrophotometer. A standard curve is obtained by plotting the absorbance (linear) versus the corresponding concentrations of the human HNP1-3 standards (log).
- The human HNP1-3 concentration of samples, which are run concurrently with the standards, can be determined from the standard curve.

5. KIT COMPONENTS AND STORAGE INSTRUCTIONS

Kit component	Cat. #	Quantity HK317-01	Quantity HK317-02	Color code
Wash buffer 40x	WB01	1 vial (30 ml)	1 vial (30 ml)	Colorless
Dilution buffer 10x	DB90	1 vial (20 ml)	1 vial (20 ml)	Green
Sample dilution buffer 10x	PD55	1 vial (15 ml)	1 vial (15 ml)	Blue
Standard		2 vials, lyophilized	4 vials, lyophilized	White
Tracer, biotinylated		1 vial, 1 ml lyophilized	2 vials, 1 ml lyophilized	White
Streptavidin-peroxidase 100x	CON03	1 tube, 0.25 ml in solution	1 tube, 0.25 ml in solution	Brown
TMB substrate	TMB050 / TMB100	1 vial (11 ml)	1 vial (22 ml)	Brown
Stop solution	STOP110	1 vial (22 ml)	1 vial (22 ml)	Red
12 Microtiter strips, pre-coated		1 plate	2 plates	
Certificate of Analysis		1	1	
Manual		1	1	
Data collection sheet		2	2	

Table 2

- Upon receipt, store individual components at 2 - 8°C. Do not freeze.
- Do not use components beyond the expiration date printed on the kit label.
- The standard and tracer in lyophilized form and the streptavidin-peroxidase in concentrated solution are stable until the expiration date indicated on the kit label, if stored at 2 - 8°C.
- The exact amount of the standard is indicated on the label of the vial and the Certificate of Analysis.
- The standard is single use. After reconstitution the standard cannot be stored.
- Once reconstituted the tracer is stable for 1 month if stored at 2 - 8°C.
- The streptavidin-peroxidase can only be stored in concentrated solution and is not stable when stored diluted.
- Upon receipt, foil pouch around the plate should be vacuum-sealed and unpunctured. Any irregularities to aforementioned conditions may influence plate performance in the assay.
- Return unused strips immediately to the foil pouch containing the desiccant pack and reseal along the entire edge of the zip-seal. Quality guaranteed for 1 month if stored at 2 - 8°C.

Materials required but not provided

- Calibrated micropipettes and disposable tips.
- Distilled or de-ionized water.
- Plate washer: automatic or manual.
- Polypropylene tubes.
- Calibrated ELISA plate reader capable of measuring absorbance at 450 nm.
- Adhesive covers can be ordered separately. Please contact your local distributor.
- Centrifuge for 1 ml tubes.

6. WARNINGS AND PRECAUTIONS

- For research use only, not for diagnostic or therapeutic use.
- This kit should only be used by qualified laboratory staff.
- Do not add under any circumstances sodium azide as preservative to any of the components.
- Do not use kit components beyond the expiration date.
- Do not mix reagents from different kits and lots. The reagents have been standardized as a unit for a given lot. Use only the reagents supplied by manufacturer.
- The assay has been optimized for the indicated standard range. Do not change the standard range.
- Open vials carefully: vials are under vacuum.
- It is advised to spin down streptavidin-peroxidase tubes before use.
- Do not ingest any of the kit components.
- Kit reagents contain 2-chloroacetamide as a preservative. 2-Chloroacetamide is harmful in contact with skin and toxic if swallowed. In case of accident or if you feel unwell, seek medical advise immediately.
- The TMB substrate is light sensitive, keep away from bright light. The solution should be colorless until use.
- The stop solution contains 2% oxalic acid and can cause irritation or burns to respiratory system, skin and eyes. Direct contact with skin and eyes should be strictly avoided. If contact occurs, rinse immediately with plenty of water and seek medical advise.
- Incubation times, incubation temperature and pipetting volumes other than those specified may give erroneous results.
- Do not reuse micro wells or pour reagents back into their bottles once dispensed.
- Handle all biological samples as potentially hazardous and capable of transmitting diseases.
- Hemolyzed, hyperlipemic, heat-treated or contaminated samples may give erroneous results.
- Use polypropylene tubes for preparation of standard and samples. Do not use polystyrene tubes or sample plates.

7. SAMPLE PREPARATION

Collection and handling

Serum and plasma samples

Please be aware that human HNP1-3 is released from neutrophils into serum during the process of blood coagulation. This will lead to false positive results. Most reliable results are obtained with heparin or EDTA plasma. In addition, it is advised to use 'careful plasma', which can be obtained as follows. Keep freshly collected blood on ice. Within 20 minutes after blood sampling, separate plasma by centrifugation: 1500xg at 4°C for 15 min. Remove plasma and transfer to fresh polypropylene tube. Be careful to not disturb white cells in the buffy coat. Recentrifuge the transferred plasma in order to avoid every contamination with white blood cells (1500xg at 4°C for 15 min).

Bronchoalveolar lavage fluid (BALF)

Perform BAL during bronchoscopy by standardized washing of the segment. Aspirate each aliquot of physiologic fluid immediately after inspiration. Discard the first aliquot of recovered BALF. Collect the BALF in polypropylene tubes and keep it on ice. Separate cells from BALF by centrifugation (500xg at 4°C for 5 min). Filter cell free BALF through a layer of gauze to remove mucus strands.

Cervicovaginal secretions samples

Place swabs in a tube containing 5 ml sterile phosphate buffered saline and immediately put on ice. Centrifuge samples at 800 g for 10 minutes at 4°C and separate supernatants from the cell pellet. Store samples at -80°C.

Feces

HNP1-3 can be measured in faeces if samples are extracted using the following extraction buffer: 0.1 M Tris, 0.15 M NaCl, 1.0 M urea, 10 mM CaCl₂, 0.1 M citric acid monohydrate, 5 g/l BSA and 0.25 mM thimerosal (pH 8.0). Add 5 ml extraction buffer to 100 mg sample (giving a dilution factor of 51, assuming the density of faeces to be 1 g/ml). Vortex samples and filter the samples to remove coarse particles (> 0.6 mm). Shake the filtrate for 20 minutes and centrifuge samples: 10,000xg at 4°C for 20 minutes. Use supernatant for analysis.

Sputum

Drink plenty of fluid the night before collection. Sit upright to collect sputum of the first cough in the morning before eating. Rinse the mouth with plain water, do not brush teeth or use mouthwash before collection of sputum. Take three good, deep breaths and cough deeply to bring up secretions (not saliva). Expectorate sputum directly into the container. Do not contaminate the rim of the container with sputum.

Storage

Store samples below -20°C, preferably at -70°C in polypropylene tubes. Storage at -20°C can affect recovery of human HNP1-3. Use samples within 24 hours after thawing. Avoid multiple freeze-thaw cycles which may cause loss of human HNP1-3 activity and give erroneous results.

Do not use hemolyzed, hyperlipemic, heat-treated or contaminated samples.

Before performing the assay, samples should be brought to room temperature (18 – 25°C) and mixed gently. Prepare all samples (controls and test samples) prior to starting the assay procedure. Avoid foaming.

Dilution procedures

HNP1-3 are highly absorbing to Ig, other proteins and plastics. Samples can only be measured accurately if diluted with supplied sample dilution buffer.

Serum and plasma samples

Human HNP1-3 can be measured accurately if plasma and serum samples are diluted at least 20x with supplied sample dilution buffer in polypropylene tubes.

Rhesus monkey plasma samples need to be diluted at least 4x for accurate measurement.

Note that most reliable results are obtained with heparin or EDTA plasma.

BALF samples

Human HNP1-3 can be measured accurately if BALF samples are diluted at least 40x with supplied sample dilution buffer in polypropylene tubes.

Feces samples

Human HNP1-3 can be measured accurately if feces samples are diluted at least 20x with supplied sample dilution buffer in polypropylene tubes.

Cervicovaginal secretions samples

Human HNP1-3 can be measured accurately if cervicovaginal secretions samples are undiluted. Further dilutions need to be prepared with supplied sample dilution buffer in polypropylene tubes.

Saliva samples

Human HNP1-3 can be measured accurately if saliva samples are diluted at least 5x with supplied sample dilution buffer in polypropylene tubes.

Sputum samples

Human HNP1-3 can be measured accurately if sputum samples are diluted at least 40x with supplied sample dilution buffer in polypropylene tubes.

Remark regarding recommended sample dilution

The mentioned dilution for samples is a minimum dilution and should be used as a guideline. The recovery of human HNP1-3 from an undiluted sample is not 100% and may vary from sample to sample. When testing less diluted samples it is advisable to run recovery experiments to determine the influence of the matrix on the detection of human HNP1-3.

Do not use polystyrene tubes or sample plates for preparation or dilution of the samples.

Guideline for dilution of samples

Please see the table on the next page for recommended sample dilutions. Volumes are based on a total volume of at least 230 µl of diluted sample, which is sufficient for one sample in duplicate in the ELISA. For dilution of samples we recommend to use at least 10 µl of sample.

	Dilution	Pre-dilution	Amount of sample or pre-dilution required	Amount of Sample Dilution buffer required
1.	10x	Not necessary	25 µl (sample)	225 µl
2.	20x	Not necessary	15 µl (sample)	285 µl
3.	50x	Not necessary	10 µl (sample)	490 µl
4.	100x	Not necessary	10 µl (sample)	990 µl
5.	500x	Recommended: 10x (see nr.1)	10 µl (pre-dilution)	490 µl
6.	1000x	Recommended: 10x (see nr.1)	10 µl (pre-dilution)	990 µl
7.	2000x	Recommended: 20x (see nr.2)	10 µl (pre-dilution)	990 µl
8.	5000x	Recommended: 50x (see nr.3)	10 µl (pre-dilution)	990 µl

Table 3

8. REAGENT PREPARATION

Allow all the reagents to equilibrate to room temperature (20 – 25°C) prior to use. Return to proper storage conditions immediately after use.

Wash buffer

Prepare wash buffer by mixing 30 ml of 40x wash buffer with 1170 ml of distilled or de-ionized water, which is sufficient for 2 x 96 tests. In case less volume is required, prepare the desired volume of wash buffer by diluting 1 part of the 40x wash buffer with 39 parts of distilled or de-ionized water.

Dilution buffer

Prepare dilution buffer by mixing 20 ml of the 10x dilution buffer with 180 ml of distilled or de-ionized water, which is sufficient for 2 x 96 tests. In case less volume is required, prepare the desired volume of dilution buffer by diluting 1 part of the 10x dilution buffer with 9 parts of distilled or de-ionized water. Concentrated dilution buffer may contain crystals. In case the crystals do not disappear at room temperature within 1 hour, concentrated dilution buffer can be warmed up to 37°C. Do not shake the solution.

Sample dilution buffer

Prepare sample dilution buffer by mixing 15 ml of the 10x sample dilution buffer with 135 ml of prepared dilution buffer, which is sufficient for 2 x 96 tests. In case less volume is required, prepare the desired volume of sample dilution buffer by diluting 1 part of the 10x plasma dilution buffer with 9 parts of dilution buffer.

Standard solution

The standard is reconstituted by pipetting the amount of sample dilution buffer mentioned on the CoA in the standard vial. Use the standard vial as Tube 1 in Figure 1. Prepare each HNP1-3 standard in polypropylene tubes by serial dilution of the reconstituted standard with Sample dilution buffer as shown in Figure 1*. After reconstitution the standard cannot be stored for repeated use.




Figure 1

*) CoA: Certificate of Analysis, Rec. St: Reconstituted Standard, Sdb: Sample Dilution buffer

Tracer solution

The tracer is reconstituted by pipetting 1 ml distilled or de-ionized water. Dilute the reconstituted 1 ml tracer with 11 ml dilution buffer, which is sufficient for 1 x 96 tests. In case

less volume is required, prepare the desired volume of tracer by diluting 1 part of the reconstituted tracer with 11 parts of dilution buffer.

Streptavidin-peroxidase solution

It is advised to spin down streptavidin-peroxidase tubes before use. Prepare the streptavidin-peroxidase solution by mixing 0.25 ml of the 100x streptavidin-peroxidase solution with 24.75 ml dilution buffer, which is sufficient for 2 x 96 tests. In case less volume is required, prepare the desired volume of streptavidin-peroxidase solution by diluting 1 part of the 100x streptavidin-peroxidase solution with 99 parts of dilution buffer.

9. ELISA PROTOCOL

Bring all reagents to room temperature (20 - 25°C) before use.

1. Determine the number of test wells required, put the necessary micro well strips into the supplied frame, and fill out the data collection sheet. Return the unused strips to the storage bag with desiccant, seal and store at 2 - 8°C.
2. Transfer 100 µl in duplicate of standard, samples, or controls into appropriate wells. Do not touch the side or bottom of the wells.
3. Cover the tray and tap the tray to eliminate any air bubbles. Be careful not to splash liquid onto the cover.
4. Incubate the strips or plate for 1 hour at room temperature.
5. Wash the plates 4 times with wash buffer using a plate washer or as follows*:
 - a. Carefully remove the plate sealer, avoid splashing.
 - b. Empty the plate by inverting plate and shaking contents out over the sink, keep inverted and tap dry on a thick layer of tissues.
 - c. Add 200 µl of wash buffer to each well, wait 20 seconds, empty the plate as described in 5b.
 - d. Repeat the washing procedure 5b/5c three times.
 - e. Empty the plate and gently tap on thick layer of tissues.
6. Add 100 µl of diluted tracer to each well using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
7. Cover the tray and incubate the tray for 1 hour at room temperature.
8. Repeat the wash procedure described in step 5.
9. Add 100 µl of diluted streptavidin-peroxidase to each well, using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
10. Cover the tray and incubate the tray for 1 hour at room temperature.
11. Repeat the wash procedure described in step 5.
12. Add 100 µl of TMB substrate to each well, using the same pipetting order as applied in step 2. Do not touch the side or bottom of the wells.
13. Cover the tray and incubate the tray for 30 minutes at room temperature. It is advised to control the reaction on the plate regularly. In case of strong development the TMB reaction can be stopped sooner. Avoid exposing the micro well strips to direct sunlight. Covering the plate with aluminium foil is recommended.
14. Stop the reaction by adding 100 µl of stop solution with the same sequence and timing as used in step 12. Mix solutions in the wells thoroughly by gently swirling the plate. Gently tap the tray to eliminate any air bubbles trapped in the wells.
15. Read the plate within 30 minutes after addition of stop solution at 450 nm using a plate reader, following the instructions provided by the instrument's manufacturer.

*) In case plate washer is used, please note: use of a plate washer can result in higher background and decrease in sensitivity. We advise validation of the plate washer with the manual procedure. Make sure the plate washer is used as specified for the manual method.

10. INTERPRETATION OF RESULTS

- Calculate the mean absorbance for each set of duplicate standards, control and samples.
- If individual absorbance values differ by more than 15% from the corresponding mean value, the result is considered suspect and the sample should be retested.
- The mean absorbance of the zero standard should be less than 0.3.
- Create a standard curve using computer software capable of generating a good curve fit. The mean absorbance for each standard concentration is plotted on the vertical (Y) axis versus the corresponding concentration on the horizontal (X) axis (logarithmic scale).
- If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.
- Samples that give a mean absorbance above the absorbance for the highest standard concentration are out of range of the assay. These samples should be retested at a higher dilution.

11. TECHNICAL HINTS

- User should be trained and familiar with ELISA assays and test procedure.
- If you are not familiar with the ELISA technique it is recommended to perform a pilot assay prior to evaluation of your samples. Perform the assay with a standard curve only following the instructions.
- Improper or insufficient washing at any stage of the procedure will result in either false positive or false negative results. Completely empty wells before dispensing wash buffer, fill with wash buffer as indicated for each cycle and do not allow wells to sit uncovered or dry for extended periods.
- Since exact conditions may vary from assay to assay, a standard curve must be established for every run. Samples should be referred to the standard curve prepared on the same plate.
- Do not mix reagents from different batches, or other reagents and strips. Remainders should not be mixed with contents of freshly opened vials.
- Each time the kit is used, fresh dilutions of standard, sample, tracer, streptavidin-peroxidase and buffers should be made.
- Caps and vials are not interchangeable. Caps should be replaced on the corresponding vials.
- To avoid cross-contaminations, change pipette tips between reagent additions of each standard, between sample additions, and between reagent additions. Also, use separate reservoirs for each reagent.
- The waste disposal should be performed according to your laboratory regulations.

Technical support

Do not hesitate to contact our technical support team at support@hycultbiotech.com for inquiries and technical support regarding the human HNP1-3 ELISA.

Hycult Biotech, Frontstraat 2A, 5405 PB Uden, the Netherlands

T: +31 (0)413 251 335, F: +31 (0)413 248 353

12. QUALITY CONTROL

The Certificate of Analysis included in this kit is lot specific and is to be used to verify results obtained by your laboratory. The absorption values provided on the Certificate of Analysis are to be used as a guideline only. The results obtained by your laboratory may differ.

This assay is designed to eliminate interference by soluble receptors, binding proteins, and other factors present in biological samples. Until all factors have been tested in the Hycult Biotech immunoassay, the possibility of interference cannot be excluded.


For optimal performance of this kit, it is advised to work according to good laboratory practice.

13. PERFORMANCE CHARACTERISTICS

Recovery

Normal extracted human fecal samples, containing low and high levels of human HNP1-3. Samples were measured using the ELISA. Values for human HNP1-3 ranged between 85.5% and 91.5%.

Linearity of sputum samples



14. TROUBLESHOOTING

Warranty claims and complaints in respect of deficiencies must be logged before expiry date of the product. A written complaint containing lot number of the product and experimental data shall be sent to support@hycultbiotech.com.

Suggestions summarized below in Table 3 can be used as guideline in case of unexpected assay results.

Low absorbance	High absorbance	Poor duplicates	All wells positive	All wells negative	Possible cause
•	•		•	•	Kit materials or reagents are contaminated or expired
•					Incorrect reagents used
•		•	•		Lyophilized reagents are not properly reconstituted
•	•	•	•	•	Incorrect dilutions or pipetting errors
•		•			Improper plastics used for preparation of standard and/or samples
•	•				Improper incubation times or temperature
		•			Especially in case of 37°C incubation: plates are not incubated uniformly
•					Assay performed before reagents were adapted to room temperature
•	•	•	•	•	Procedure not followed correctly
				•	Omission of a reagent or a step
		•			Poor mixing of samples
	•		•		Low purity of water
	•	•			Strips were kept dry for too long during/after washing
	•	•	•		Inefficient washing
	•	•			Cross-contamination from other samples or positive control
		•	•		TMB solution is not clear or colorless
•	•				Wrong filter in the micro titer reader
	•	•			Air bubbles
		•			Imprecise sealing of the plate after use
•					Wrong storage conditions
•					Lamp in microplate reader is not functioning optimally

Table 4

15. REFERENCES

1. Zhang, L et al; Contribution of human alpha-defensin-1, -2 and -3 to the anti-HIV-1 activity of CD8 antiviral factor. *Science* 2002, *298*: 995
2. Espinoza, J et al; Antimicrobial peptides in amniotic fluid: defensins, calprotectin and bacterial/permeability-increasing protein in patients with microbial invasion of the amniotic cavity, intra-amniotic inflammation, preterm labor and premature rupture of membranes. *J Matern Fetal Neonatal Med* 2003, *13*: 2
3. Nelsestuen, G et al; Proteomic identification of human neutrophil alpha-defensins in chronic lung allograft rejection. *Proteomics* 2005, *5*: 1705
4. Baroncelli, S et al; Characterization of α -defensins plasma levels in *Macaca fascicularis* and correlations with virological parameters during SHIV89.6Pcy11 experimental infection. *AIDS Res Hum Retroviruses* 2007, *23*: 287
5. Levinson, P et al; HIV-Neutralizing Activity of Cationic Polypeptides in Cervicovaginal Secretions of Women in HIV Serodiscordant Relationships. *PlosOne* 2012, *7*:e31996
6. Eick, S et al; Lack of cathelicidin processing in Papillon-Lefèvre syndrome patients reveals essential role of LL-37 in periodontal homeostasis. *Orphanet Journal of Rare Diseases* 2014, *9*:148
7. Rohde, G et al; CXC chemokines and antimicrobial peptides in rhinovirus-induced experimental asthma exacerbations. *Clinical & Experimental Allergy* 2014, *44*:930
8. O'Donnell, L et al; The Oral Microbiome of Denture Wearers Is Influenced by Levels of Natural Dentition. *PLos One* 2015, *10*: e0137717
9. Thijs, W et al; Nasal Levels of Antimicrobial Peptides in Allergic Asthma Patients and Healthy Controls: Differences and Effect of a Short 1,25(OH)₂ Vitamin D3 Treatment. *Plos One* 2015, *10*: e0140986
10. Qi, Y et al; Elevated Plasma α -Defensins (HNP1–3) Levels Correlated with IgA1 Glycosylation and Susceptibility to IgA Nephropathy. *Disease markers* 2016