

CERTIFICATE OF ANALYSIS – TECHNICAL DATA SHEET

Product name	ICAM-1, Human, pAb		
Catalog number	HP9018		
Lot number	-	Expiry date	-
Volume	1 ml	Amount	100 µg
Formulation	0.2 µm filtered in PBS+0.1%BSA+0.02%NaN3	Concentration	100 µg/ml
Host Species	Rabbit IgG	Conjugate	None
Endotoxin	N.A.	Purification	Protein G
Storage	4°C		

Application notes

	IHC-F	IHC-P	IF	FC	FS	IA	IP	W
Reference #								
Yes	•					•	•	•
No								
N.D.		•	•	•	•			

N.D.= Not Determined; IHC = Immuno histochemistry; F = Frozen sections; P = Paraffin sections; IF = Immuno Fluorescence; FC = Flow Cytometry; FS = Functional Studies; IA = Immuno Assays; IP = Immuno Precipitation; W = Western blot

Dilutions to be used depend on detection system applied. It is recommended that users test the reagent and determine their own optimal dilutions. The typical starting working dilution is 1:10.

- IHC-F: For selection of the most useful dilution in a given situation a test staining with cells or tissue known to express the antigen should be performed. To this end a small fresh skin biopsy can be incubated for 4 hours with TNF-alpha (1 ng/ml), IL-1 (100 U/ml) or LPS (1 µg/ml) in tissue culture medium at 37°C.

General Information

Description	The antibody reacts with both intact human ICAM-1, CD54 and with soluble human ICAM-1. The antibody reacts also with ICAM-1 present on chimpanzee, rhesus monkey, cynomolgus monkey and baboon tissues.
Aliases	CD54
Cross reactivity	Chimpanzee: Yes; Rhesus monkey: Yes; Cynomolgous monkey: Yes; Baboon: Yes
Storage&stability	Product should be stored at 4°C. Under recommended storage conditions, product is stable for at least one year.
Precautions	For research use only. Not for use in or on humans or animals or for diagnostics. It is the responsibility of the user to comply with all local/state and federal rules in the use of this product. Hycult Biotech is not responsible for any patent infringements that might result from the use or derivation of this product.

We hereby certify that the above-stated information is correct and that this product has been successfully tested by the Quality Control Department. This product was released for sale according to the existing specifications. This document has been produced electronically and is valid without a signature.

Approved by Manager of QC
Robbert Zwinkels

Date
13/03/2018

Do you have any questions or comments regarding this product? Please contact us via support@hycultbiotech.com.